

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Lens Removal & Anterior Vitrectomy System
Name and address of the applicant	Med-Logics, Inc. 1627 Enterprise St Athens, TX 75751 United States of America
Name and address of the manufacturer	Med-Logics, Inc. 1627 Enterprise St Athens, TX 75751 United States of America
Name and address of the factory	<input type="checkbox"/> Additional information on page 2 Med-Logics, Inc. 1627 Enterprise St Athens, TX 75751 United States of America
Note: When more than one factory, please report on page 2	
Ratings and principal characteristics	100-240 VAC, 50/60 Hz, 5A
Trademark (if any)	CataPulse
Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	CP1700 CataPulse Console, CP1800 CataPulse Footswitch and CP1900 CataPulse Cart
Additional information (if necessary may also be reported on page 2)	<input type="checkbox"/> Additional information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1-6:2010/AMD1:2013, IEC 60601-1:2005/AMD1:2012 National differences: EU Group Differences, EU Special National Conditions, EU A-Deviations, CA, CH, JP, KR, US Comments: Also tested to be in conformity with IEC 80601-2-58: 2014, report no. 102945703LAX-004. Also evaluated to IEC 62304:2006/AMD1:2015, report no. 102945703LAX-003.
As shown in the Test Report Ref. No. which forms part of this Certificate	102945703LAX-001, 102945703LAX-002

This CB Test Certificate is issued by the National Certification Body

Intertek Testing Services N.A., Inc.
545 E. Algonquin Road, Suite F, Arlington Heights, IL 60005, United States of America